

Ophthalmic OT Table



Technical Constraints

Length of Table	2000mm
Width of Table Top	540mm
Minimum Height Without Mattress	Min.640mm
Maximum Height Without Mattress	Max.940mm
Leg section	0°-90°

OPTIONAL ACCESSORIES



**Foot Switch
(Optional)**



**Instrument Tray
(Optional)**

STANDARD ACCESSORIES




Knee/Lithomic Crutches



Lateral Support



Anaesthesia Screen



Arm Board



Shoulder Support

For Further Information, contact:

**# 36 B, Bajrang Bali Colony, Rohtak Gate,
Bhiwani - 127021 (HR)**

M.: 9802324555, 8826707666

Email: nvsurgicals666@gmail.com